

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 1/6/2015

AMS
Jul

Instrument Serial Number: 010500	As of 06-Jan-15
----------------------------------	-----------------

Start Date	End Date	Agency Name
06-Jan-15		Franklin PD
17-Dec-14	06-Jan-15	DFS Central Lab
16-Jul-14	17-Dec-14	Loudoun County SO
18-Jun-14	16-Jul-14	DFS Central Lab
02-Dec-08	18-Jun-14	Martinsville PD
19-May-08	02-Dec-08	DFS Central Lab

Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/6/2014 To 1/6/2015

Handwritten initials/signature

Instrument Serial Number: 010500 As of 06-Jan-15

Maintenance Date	Certification Date	Technician Initials	Remarks
06-Jan-15	23-Dec-14	HMS (18910)	REPLACED UPS BACK-UP BATTERY.
23-Dec-14	23-Dec-14	HMS (18910)	CERTIFIED
10-Jul-14	10-Jul-14	DBH (21250)	CERTIFIED.
08-Jan-14	08-Jan-14	DBH (21250)	CERTIFIED. REPLACED DRY GAS STANDARD.

INTOX EC/IR II
Quality Assurance Worksheet

ANS
sid

Instrument Serial Number

010500

Worksheet Start Date

1/6/2015

Location

Franklin Police Department

Address

1018 Pretlow St., Franklin, VA 23851

DFS Technician

Heather Stanton

License No.

18910

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist.

☒

Instrument Barometer (mm HG)

767

Reference Barometer (mm HG)

768

Reference Barometer(RB)Serial #

009113

RB Calibration Due

4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.309
0.300	0.303	0.293	0.312	Sample 2	0.307
Precision		sample min	sample max	Sample 3	0.307
0.002		0.307	0.309		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.101	0.097	0.104	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		

Dry gas standard Lot No. (with tank no.)

AG332301-18

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies

Mouthpieces

Certificates of Analysis

Operator Worksheet

Other:

50

50

2 pads

Notes:

Replaced UPS back-up battery.

Instrument Serial Number

010500

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath n Stock

Date

1/6/15

Issuing Analyst

jud

Date

1/12/15

hms
jld

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010500 Test Number: 565

Test Date: 01/06/2015 Test Time: 11:00 EST

Dry Gas Target: 0.303

Lot Number: AG428003-09 Exp Date: 10/07/2016

Tank Pressure: 502 psi Barometric Pressure: 767 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:01
CHK	0.309	11:01
BLK	0.000	11:03
CHK	0.307	11:04
BLK	0.000	11:06
CHK	0.307	11:07

Test Status: *Success*

Calibration CRC: 619E4344

AMS
Jed

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010500 Test Number: 566

Test Date: 01/06/2015 Test Time: 11:09 EST

Dry Gas Target: 0.101

Lot Number: AG332301-18 Exp Date: 11/19/2015


Tank Pressure: 517 psi Barometric Pressure: 766 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:10
CHK	0.100	11:11
BLK	0.000	11:12
CHK	0.100	11:13
BLK	0.000	11:14
CHK	0.100	11:15

Test Status: *Success*

Calibration CRC: 619E4344



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 01/06/2015
TEST EQUIPMENT NUMBER 010500		

RESULTS: TIME SAMPLE TAKEN **11:23 EST**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 06 Jan 2015 at 15:59

Handwritten signature/initials

Test Results

Instrument Serial Number 010500

Test # 000567 Subject Test

Test Location 1 Department of
Test Date 06 Jan 2015

Test Location 2 Forensic Science
Test Time 11:16
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 06 Jan 2015 End Time 11:24

Result Time 11:23

Result Date 06 Jan 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:17

Data Type BLK

Sample Value 0.000

Sample Time 11:18

Data Type CHK

Sample Value 0.100

Sample Time 11:18

Data Type BLK

Sample Value 0.000

Sample Time 11:20

Data Type SUBJ

Sample Value 0.000

Sample Time 11:20

Data Type BLK

Sample Value 0.000

Sample Time 11:21

Data Type SUBJ

Sample Value 0.000

Sample Time 11:23

Data Type BLK

Sample Value 0.000

Sample Time 11:24

Standard Type Dry Gas Std

Standard Value 0.101

Standard Lot Number AG332301-18

Standard Expiration Date 11/19/2015

Tank Pressure 512

Barometric Pressure 766 mmHg

Blow Sample Number 1 Blow Duration 3.34 sec

Blow Volume 1789 cc End-of-Blow Time 11:20

Blow Sample Number 2 Blow Duration 3.45 sec

Blow Volume 1773 cc End-of-Blow Time 11:23

Tamper Evident Stamp d7092e48

Test Status Code 0

Test Status Success